

Depth Sensor Verification Log - DEP Regional Operation Centers

SOP - Status and Trend Networks Sampling Manual and ROC Training Manual.

Report two decimal places for electronic devices. Report one decimal place for manual devices. Numbers ≤ 4, are rounded down; numbers ≥ 5 are rounded up.

Acceptance criteria for electronic device verification ± 5%. Acceptance criteria for analog device incremental markings check ± 10%. Acceptance criteria for analog device total length of line verification ± 5%.

Time Zone (circle one): ETZ / CTZ

Field Device Type (Field Meter / Sonar / Secchi)	Field Device ID	Field Meter Serial Number	Verification Location	Reference Device	Verification Date	Verification Time	Electronic Device Verification - Field Device Depth (m)	Electronic Device Verification - Reference Device Depth (m)	Electronic Device Verification - Result (circle one)	Analog Device - Incremental Markings Check (every 0.1m) - Result (circle one)	Analog Device - Total Length of Line Verification - Field Device Measurement (m)	Analog Device - Total Length of Line Verification - Reference Device Measurement (m)	Analog Device - Total Length of Line Verification - Result (circle one)	Analog Device - Markings Redone (circle one)	Activity Performed By (Staff Name)	Comments
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		
									P / F	P / F			P / F	Y / N		