

Quarterly Temperature Verification Log - DEP Regional Operation Centers

DEP SOP FT 1400. Acceptance Criteria for Temp. $\pm 0.5^{\circ}\text{C}$.

Record all digits displayed for temperature readings. Do not round before reporting measurements.

CCV = Continuing Calibration Verification. Target temperature for cold CCV is 0 - 10 ($^{\circ}\text{C}$). Target temperature for hot CCV is 30 - 40 ($^{\circ}\text{C}$).

Time Zone (circle one): **ETZ / CTZ**

Field Meter ID	Field Meter Serial Number	NIST Reference Device ID	Activity Date	Cold CCV Time	Cold CCV Field Meter Temp ($^{\circ}\text{C}$)	Cold CCV NIST Temp ($^{\circ}\text{C}$)	Cold CCV Result (circle one)	Hot CCV Time	Hot CCV Field Meter Temp ($^{\circ}\text{C}$)	Hot CCV NIST Temp ($^{\circ}\text{C}$)	Hot CCV Result (circle one)	Activity Performed By (Staff Name)	Comments
							P / F				P / F		
							P / F				P / F		
							P / F				P / F		
							P / F				P / F		
							P / F				P / F		
							P / F				P / F		
							P / F				P / F		
							P / F				P / F		